Application Note: Calibration Error Calculation in AirVision Calibration Results Report

The determination of % error (sometimes called % difference in EPA documents, although there is a somewhat pedantic difference) is generally familiar to all as:

$$\% \ Error = \left|\frac{T-E}{T}\right| \bullet 100$$

Where T = the "test"/expected/reference value and E = the analyzer response

However, some questions come into play when displaying values in calibration reports, and when users try to compare hand calculations from the report value and can come up with a slightly different % error calculation. The reason for this can be traced to the fact that AirVision stores a full precision value in the calibration results database, both for the analyzer response and the expected value. Example:

🚯 🖶 🙆 =			AirVision	-	
Home View Favorit	calibration Data Editor				
Retrieve Add Export Data to Excel					
Calibration Editor Options					
🚯 Data Source Details 🚯 Instrument Poli Commands 🐧 Manual Instrument Poli 🕃 Log Viewer 📊 Average Data Editor 🌾 Calibration Data Editor 🗴					
+ Selection					
Calibration Data					
Site △ Parameter △ Seque	nce Phase St	Start Date \triangle End Date \triangle	Value Expected Exclude From Value Reporting	Tracer ID Level	Method Coc
			8 8 8		
Alban O3 O3_SZ	Span_M 12/01	1/2021 01:00 12/01/2021 01:3	190.4972 200		047
<					
User: Admin Profile: LAPTOP-1SA681RS Version: 44.60 B Albany: 03 CT Phone Com					

If we use the pure floating point value, the % error is -4.75%, but if we apply EPA truncation to the values (using a value of '190' for the analyzer response), the value is -5.0%.

Which approach is correct? At this point, it seems to be an agency preference as it's not defined in the CFR, and probably embedded in varying SOPs. So, it is a configurable setting in AirVision.

The AirVision default for most systems is to use the full precision value, but to report values based on the rounding and truncation rules in the parameter (in this case, 1 digit):



A hand calculation here would still show 4.6%, and cause some consternation. The user has a few options:

- 1. Understand and agree with the "full precision" method and know that a hand calculation purely from the Calibration Results report won't agree (and to use the Calibration Data Editor to see or export full precision for auditing).
- 2. Increase the "Cal Precision" setting in the parameter editor to get closer (e.g., an alternate report precision just for cal reports):



3. Finally, an internal database preference settings can be toggled to just use the rounded / truncated value:

UPDATE UtilitySetting set SettingValue = TRUE where SettingKey = 'EDASCalErrorMethod' UPDATE UtilitySetting set SettingValue = TRUE where SettingKey = 'UsePrecForCalExpectedFormat'

As you can see from the name, the preference setting will match the calculation method from the old E-DAS Ambient application, if that is the agency preference / SOP.